


LABORATORIO PREGEN  
 Carrera 15a No 106-42  
 BOGOTA

Batch 1994, Rack A, Plate 1, Well A05, 482035  
 [8827BD96F51A2642] Nov 25, 2025 13:07:30 Pressure = 69 bar (66 to 70)

AF



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.393	F 0.20	3.2%	58274	
2	0.542	F 0.28	4.6%	83663	
3	0.893	F 0.46	5.7%	105105	
4	1.083	F 0.55	4.3%	78669	
5	1.932	F 0.99	7.3%	134594	6
6	1.988	F 1.02	12.5%	229800	Consistent with F
7	2.620	F 1.34	1.5%	26857	
8	2.760	F 1.41	2.2%	40409	
9	3.042	A 0.81	1.0%	18303	
10	3.332	A 0.89	0.2%	4547	
11	3.748	A 1.00	55.9%	1025809	A peak - REVIEW
12	4.308	A 1.14	0.3%	6027	
13	4.530	S 0.86	1.3%	24255	3
Total Area:				1836312	

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range