


LABORATORIO PREGEN  
Carrera 15a No 106-42  
BOGOTA

Batch 2098, Rack A, Plate 1, Well C04, 498831  
[89723F0156B9DB95] Mar 19, 2026 12:08:27 Pressure = 61 bar (60 to 62)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.417	F 0.48	0.9%	9823	
2	0.513	F 0.59	0.7%	7847	
3	0.722	F 0.83	13.9%	146301	Acetylated F peak
4	0.863	F 0.99	65.9%	694929	Consistent with F
5	1.252	A 1.02	18.6%	195817	A peak
Total Area:				1054717	

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range