


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2191, Rack A, Plate 1, Well H07, 493044
[D97885115E779B95] Jul 06, 2026 14:48:27 Pressure = 79 bar (78 to 81)

FA



PEAK	RT	REL	RT	% CONC	AREA	COMMENT
1	0.515	F	0.62	1.1%	12753	
2	0.682	F	0.82	12.8%	150209	Acetylated F peak
3	0.820	F	0.98	59.1%	693289	Consistent with F
4	1.173	A	1.01	27.0%	316587	A peak - REVIEW
Total Area: 1172838						

Codes: 1) Wide A peak
2) Area of A peak < 80%
3) Peak area greater than expected
4) Peak after A2
5) A1c > 10%
6) HbF or variant present
7) Total sample area too small/big
8) A2 is not within normal range