


LABORATORIO PREGEN
 Carrera 15a No 106-42
 BOGOTA

Batch 2049, Rack A, Plate 1, Well D04, 488998
 [9D72DD0AC19D51BF] Jan 20, 2026 12:19:56 Pressure = 81 bar (79 to 82)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.412	F 0.49	1.3%	11085	
2	0.498	F 0.59	0.6%	5425	
3	0.697	F 0.82	11.7%	103878	Acetylated F peak
4	0.840	F 0.99	49.6%	439166	Consistent with F
5	0.993	F 1.18	3.0%	26768	
6	1.203	A 1.01	33.7%	298571	A peak - REVIEW
Total Area:				884893	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range