


LABORATORIO PREGEN
 Carrera 15a No 106-42
 BOGOTA

Batch 2107, Rack A, Plate 1, Well B08, 498492
 [D8270494085604E2] Mar 25, 2026 11:50:06 Pressure = 57 bar (56 to 59)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.417	F 0.48	1.7%	13611	
2	0.510	F 0.58	0.9%	7116	
3	0.715	F 0.82	15.8%	130156	Acetylated F peak
4	0.867	F 0.99	58.5%	481940	Consistent with F
5	1.013	F 1.16	3.9%	31870	
6	1.265	A 1.02	19.4%	159579	A peak
Total Area:				824272	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range