


LABORATORIO PREGEN
 Carrera 15a No 106-42
 BOGOTA

Batch 2068, Rack A, Plate 1, Well C08, 498919
 [CD72606E6CFBD317] Feb 10, 2026 11:28:18 Pressure = 94 bar (94 to 97)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.413	F 0.49	0.7%	7143	
2	0.508	F 0.60	0.2%	2389	
3	0.688	F 0.82	15.3%	164493	Acetylated F peak
4	0.827	F 0.98	73.9%	793924	Consistent with F
5	1.195	A 1.01	9.8%	105733	A peak
Total Area:				1073682	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range