


- - - - -
`LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2197, Rack A, Plate 1, Well C07, 487153
[9D78C6CAEAA0446A] Jul 15, 2026 13:19:32 Pressure = 84 bar (83 to 85)

FA



PEAK	RT	REL	RT	% CONC	AREA	COMMENT
1	0.507	F	0.61	4.4%	46025	
2	0.825	F	1.00	70.6%	731176	Consistent with F
3	1.158	A	1.00	25.0%	258882	A peak - REVIEW
Total Area: 1036083						

Codes: 1) Wide A peak
2) Area of A peak < 80%
3) Peak area greater than expected
4) Peak after A2
5) A1c > 10%
6) HbF or variant present
7) Total sample area too small/big
8) A2 is not within normal range