


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2022, Rack A, Plate 1, Well D10, 482097
[8C6784D320F5D9BD] Dec 19, 2025 11:38:53 Pressure = 76 bar (75 to 77)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.403	F 0.47	1.7%	17676	
2	0.495	F 0.58	1.0%	10990	
3	0.715	F 0.83	14.4%	152809	Acetylated F peak
4	0.862	F 1.01	56.0%	596298	Consistent with F
5	0.998	F 1.17	1.2%	12829	
6	1.220	A 1.01	25.7%	273850	A peak - REVIEW

Total Area: 1064452

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range