


LABORATORIO PREGEN  
Carrera 15a No 106-42  
BOGOTA

Batch 2034, Rack A, Plate 1, Well C06, 171731  
[89639656548204C0] Jan 06, 2026 11:37:15 Pressure = 82 bar (82 to 85)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.413	F 0.48	1.4%	9685	
2	0.500	F 0.58	1.1%	7577	
3	0.707	F 0.82	16.1%	109632	Acetylated F peak
4	0.835	F 0.97	57.9%	394306	Consistent with F
5	0.987	F 1.15	3.3%	22701	
6	1.203	A 1.01	20.2%	137486	A peak - REVIEW
Total Area:				681387	

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range