


LABORATORIO PREGEN  
 Carrera 15a No 106-42  
 BOGOTA

Batch 2186, Rack A, Plate 1, Well G10, 493530-MMX  
 [D93CBA46DC10ECE0] Jun 25, 2026 15:47:27 Pressure = 74 bar (72 to 74)

FAS



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.420	F 0.23	4.0%	83361	
2	0.550	F 0.30	4.8%	98695	
3	0.857	F 0.46	22.7%	468873	
4	1.758	F 0.95	59.7%	1232213	Consistent with F
5	2.548	F 1.37	2.6%	53773	
6	3.550	A 1.00	3.6%	74110	A peak
7	5.035	S 0.98	2.5%	51410	Consistent with S
Total Area:				2062435	<b>A/V=1.44</b>

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range