


LABORATORIO PREGEN  
 Carrera 15a No 106-42  
 BOGOTA

Batch 2179, Rack A, Plate 1, Well D08, 502190  
 [CC6CAF942A1AECE8] Jun 18, 2026 14:02:53 Pressure = 70 bar (70 to 72)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.523	F 0.62	0.7%	26658	
2	0.688	F 0.82	15.6%	579886	Acetylated F peak
3	0.812	F 0.97	66.8%	2483604	Consistent with F
4	1.187	A 1.00	16.8%	626032	A peak
Total Area:				3716180	

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range