


LABORATORIO PREGEN
 Carrera 15a No 106-42
 BOGOTA

Batch 2089, Rack A, Plate 1, Well B09, 172593
 [8C76386B1B935197] Mar 06, 2026 12:09:14 Pressure = 56 bar (56 to 58)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.412	F 0.47	1.2%	15733	
2	0.512	F 0.58	0.7%	9437	
3	0.743	F 0.85	14.8%	195132	Acetylated F peak
4	0.863	F 0.99	67.8%	892504	Consistent with F
5	1.262	A 1.01	15.4%	203060	A peak
Total Area:				1315866	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range