


LABORATORIO PREGEN  
Carrera 15a No 106-42  
BOGOTA

Batch 2179, Rack A, Plate 1, Well G03, 486010  
[D82DFA94281A4E60] Jun 18, 2026 15:07:35 Pressure = 71 bar (70 to 72)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.425	F 0.51	1.0%	8931	
2	0.510	F 0.61	0.3%	2742	
3	0.680	F 0.81	14.2%	128295	Acetylated F peak
4	0.825	F 0.98	67.1%	606843	Consistent with F
5	1.200	A 1.01	17.4%	157744	A peak
Total Area:				904555	

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range