


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 1967, Rack A, Plate 1, Well A02, 170874
[9937F57F9B27733F] Nov 05, 2025 12:16:20 Pressure = 66 bar (66 to 68)

FA



| PEAK | RT | REL RT | % CONC | AREA | COMMENT |
|-------------|-------|--------|--------|--------|-------------------|
| 1 | 0.397 | F 0.46 | 0.9% | 7149 | |
| 2 | 0.502 | F 0.58 | 0.8% | 5972 | |
| 3 | 0.718 | F 0.83 | 15.8% | 125521 | Acetylated F peak |
| 4 | 0.857 | F 0.99 | 69.5% | 553053 | Consistent with F |
| 5 | 1.238 | A 1.00 | 13.0% | 103792 | A peak |
| Total Area: | | | | 795487 | |

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range