


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2194, Rack A, Plate 1, Well C07, 486298
[D968811FBED7931F] Jul 09, 2026 11:21:52

Pressure = 81 bar (81 to 83)

FA



PEAK	RT	REL	RT	% CONC	AREA	COMMENT
1	0.530	F	0.64	0.7%	8169	
2	0.688	F	0.83	11.7%	128638	Acetylated F peak
3	0.820	F	0.99	64.2%	707382	Consistent with F
4	1.148	A	1.00	23.4%	258220	A peak - REVIEW
Total Area: 1102409						

Codes: 1) Wide A peak
2) Area of A peak < 80%
3) Peak area greater than expected
4) Peak after A2
5) A1c > 10%
6) HbF or variant present
7) Total sample area too small/big
8) A2 is not within normal range