


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2018, Rack A, Plate 1, Well F10, 171503
[9823C4F405A88CEA] Dec 15, 2025 11:55:56 Pressure = 72 bar (72 to 74)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.398	F 0.46	1.0%	11712	
2	0.495	F 0.58	0.5%	5159	
3	0.748	F 0.87	11.6%	129485	Acetylated F peak
4	0.847	F 0.99	65.8%	736760	Consistent with F
5	1.012	F 1.18	0.6%	6374	
6	1.233	A 1.01	20.6%	230116	A peak - REVIEW
Total Area:				1119606	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range