


LABORATORIO PREGEN  
Carrera 15a No 106-42  
BOGOTA

Batch 2034, Rack A, Plate 1, Well F09, 171765  
[D87286132BFD53B7] Jan 06, 2026 12:58:06 Pressure = 82 bar (82 to 84)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.400	F 0.47	1.6%	11490	
2	0.490	F 0.58	0.7%	4893	
3	0.732	F 0.86	11.7%	84414	Acetylated F peak
4	0.837	F 0.98	57.5%	414669	Consistent with F
5	0.998	F 1.17	0.9%	6160	
6	1.208	A 1.02	27.4%	197265	A peak - REVIEW
7	1.382	S 0.89	0.3%	2311	

Total Area: 721202

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range