


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2051, Rack A, Plate 1, Well A01, 171873-P
[CC639D1AE995F3BF] Jan 20, 2026 14:47:56 Pressure = 79 bar (79 to 81)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.400	F 0.47	1.6%	14885	
2	0.507	F 0.60	0.3%	2993	
3	0.700	F 0.82	12.1%	111980	Acetylated F peak
4	0.845	F 0.99	68.0%	631272	Consistent with F
5	1.205	A 1.01	18.0%	167488	A peak
Total Area:				928618	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range