


LABORATORIO PREGEN  
Carrera 15a No 106-42  
BOGOTA

Batch 2077, Rack A, Plate 1, Well D05, 489342  
[8D3623D2F77B53B5] Feb 18, 2026 11:09:19 Pressure = 56 bar (55 to 57)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.422	F 0.48	0.9%	14977	
2	0.512	F 0.58	1.1%	17604	
3	0.725	F 0.83	11.2%	180128	Acetylated F peak
4	0.865	F 0.98	66.7%	1072277	Consistent with F
5	1.263	A 1.00	20.1%	323192	A peak - REVIEW
Total Area: 1608178					

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range