


LABORATORIO PREGEN  
Carrera 15a No 106-42  
BOGOTA

Batch 2149, Rack A, Plate 1, Well G07, 173440  
[CC79B01C1C5A4E68] May 20, 2026 14:48:21 Pressure = 64 bar (63 to 66)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.433	F 0.50	0.5%	6323	
2	0.522	F 0.60	0.4%	4830	
3	0.710	F 0.82	13.6%	168387	Acetylated F peak
4	0.847	F 0.98	69.5%	860852	Consistent with F
5	1.218	A 1.01	16.0%	198460	A peak
Total Area:				1238852	

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range