


LABORATORIO PREGEN  
Carrera 15a No 106-42  
BOGOTA

Batch 2121, Rack A, Plate 1, Well D02, 502315  
[D83212165CA37BB5] Apr 13, 2026 11:48:15 Pressure = 57 bar (57 to 59)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.510	F 0.58	1.2%	12148	
2	0.743	F 0.85	14.3%	148607	Acetylated F peak
3	0.858	F 0.98	59.5%	616557	Consistent with F
4	1.245	A 1.01	25.0%	258510	A peak - REVIEW
Total Area: 1035822					

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range