


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2197, Rack A, Plate 1, Well G10, 493294
[8D38C78AE220EC62] Jul 15, 2026 15:04:38

Pressure = 83 bar (83 to 86)

FA



PEAK	RT	REL	RT	% CONC	AREA	COMMENT
1	0.513	F	0.62	1.3%	14623	
2	0.702	F	0.84	11.8%	129673	Acetylated F peak
3	0.817	F	0.98	76.0%	834168	Consistent with F
4	1.158	A	1.01	10.9%	119709	A peak
Total Area: 1098173						

Codes: 1) Wide A peak
2) Area of A peak < 80%
3) Peak area greater than expected
4) Peak after A2
5) A1c > 10%
6) HbF or variant present
7) Total sample area too small/big
8) A2 is not within normal range