


LABORATORIO PREGEN  
Carrera 15a No 106-42  
BOGOTA

Batch 1364, Rack A, Plate 1, Well C10, 27448-P  
[888DB380080F373D] Apr 08, 2024 10:42:29 Pressure = 61 bar (61 to 62)

FA



| PEAK        | RT    | REL RT | % CONC | AREA    | COMMENT           |
|-------------|-------|--------|--------|---------|-------------------|
| 1           | 0.408 | F 0.46 | 2.2%   | 30542   |                   |
| 2           | 0.485 | F 0.54 | 4.1%   | 57050   |                   |
| 3           | 0.867 | F 0.97 | 87.5%  | 1210765 | Consistent with F |
| 4           | 1.258 | A 1.01 | 6.1%   | 84936   | A peak            |
| Total Area: |       |        |        | 1383293 |                   |

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range