


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2197, Rack A, Plate 1, Well F10, 487159
[DC69C2DFB7571335] Jul 15, 2026 14:40:23 Pressure = 84 bar (84 to 86)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.518	F 0.62	0.8%	13244	
2	0.683	F 0.82	11.0%	188282	Acetylated F peak
3	0.817	F 0.98	73.6%	1257899	Consistent with F
4	1.152	A 1.00	14.6%	250206	A peak
Total Area: 1709631					

Codes: 1) Wide A peak
2) Area of A peak < 80%
3) Peak area greater than expected
4) Peak after A2
5) A1c > 10%
6) HbF or variant present
7) Total sample area too small/big
8) A2 is not within normal range