


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2182, Rack A, Plate 1, Well G10, 493539
[9D2DEE68BBCFB33F]

Jun 23, 2026 12:56:52

Pressure = 71 bar (71 to 73)

FA



PEAK	RT	REL	RT	% CONC	AREA	COMMENT
1	0.422	F	0.50	1.3%	14084	
2	0.510	F	0.61	0.6%	6074	
3	0.698	F	0.83	12.8%	138803	Acetylated F peak
4	0.830	F	0.99	64.5%	697463	Consistent with F
5	1.180	A	1.01	20.8%	224345	A peak - REVIEW
Total Area: 1080769						

Codes: 1) Wide A peak
2) Area of A peak < 80%
3) Peak area greater than expected
4) Peak after A2
5) A1c > 10%
6) HbF or variant present
7) Total sample area too small/big
8) A2 is not within normal range