


LABORATORIO PREGEN
 Carrera 15a No 106-42
 BOGOTA

Batch 2077, Rack A, Plate 1, Well C01, 492105
 [DC377792FDFB531D] Feb 18, 2026 10:34:57 Pressure = 56 bar (56 to 58)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.412	F 0.47	0.8%	10194	
2	0.517	F 0.59	0.7%	9187	
3	0.725	F 0.83	10.8%	142510	Acetylated F peak
4	0.867	F 0.99	62.6%	829228	Consistent with F
5	1.265	A 1.02	25.2%	333089	A peak - REVIEW
Total Area:				1324208	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range