


LABORATORIO PREGEN  
Carrera 15a No 106-42  
BOGOTA

Batch 2065, Rack A, Plate 1, Well B04, 489200  
[D97230FB30F9591D] Feb 04, 2026 12:08:33 Pressure = 86 bar (85 to 88)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.417	F 0.50	0.7%	7090	
2	0.503	F 0.60	0.4%	4342	
3	0.698	F 0.84	15.0%	146091	Acetylated F peak
4	0.837	F 1.00	70.5%	688506	Consistent with F
5	1.190	A 1.01	13.4%	130896	A peak
Total Area:				976925	

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range