


LABORATORIO PREGEN
 Carrera 15a No 106-42
 BOGOTA

Batch 1850, Rack A, Plate 1, Well C06, 785261 FA
 [8974C8867DB726C2] Jul 02, 2025 11:29:30 Pressure = 61 bar (60 to 62)



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.402	F 0.46	1.2%	15585	
2	0.490	F 0.57	0.9%	12033	
3	0.700	F 0.81	12.9%	170853	Acetylated F peak
4	0.838	F 0.97	59.2%	785433	Consistent with F
5	0.993	F 1.15	1.2%	15511	
6	1.225	A 1.02	24.4%	323521	A peak - REVIEW
7	1.377	A 1.14	0.3%	4463	
Total Area:				1327399	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range