


LABORATORIO PREGEN
 Carrera 15a No 106-42
 BOGOTA

Batch 1827, Rack A, Plate 1, Well B08, 455899 FA
 [9C64C5778CDF0640] Jun 11, 2025 09:52:19 Pressure = 59 bar (59 to 61)



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.400	F 0.47	1.3%	18141	
2	0.493	F 0.58	0.9%	12275	
3	0.718	F 0.85	15.8%	222931	Acetylated F peak
4	0.840	F 0.99	70.8%	999494	Consistent with F
5	1.225	A 1.01	11.2%	158013	A peak
Total Area:				1410854	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range