


LABORATORIO PREGEN  
Carrera 15a No 106-42  
BOGOTA

Batch 2008, Rack A, Plate 1, Well F06, 484349  
[DD73EA242638AE6A] Dec 09, 2025 13:09:08 Pressure = 72 bar (72 to 74)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.500	F 0.58	2.3%	36342	
2	0.705	F 0.82	16.6%	259835	Acetylated F peak
3	0.843	F 0.97	57.7%	903858	Consistent with F
4	0.998	F 1.15	1.8%	28850	
5	1.230	A 1.01	21.5%	336466	A peak - REVIEW

Total Area: 1565351

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range