


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 1964, Rack A, Plate 1, Well E10, 482292
[8C67A590A0502E68] Oct 31, 2025 12:35:01 Pressure = 65 bar (65 to 67)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.405	F 0.47	1.0%	13526	
2	0.495	F 0.57	0.7%	10464	
3	0.712	F 0.82	5.7%	80730	
4	0.853	F 0.99	73.9%	1050772	Consistent with F
5	1.243	A 1.00	18.7%	266616	A peak
Total Area:				1422108	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range