


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2068, Rack A, Plate 1, Well H06, 172249
[CD63307B318C0EE2] Feb 10, 2026 13:29:34 Pressure = 94 bar (94 to 97)

AF



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.395	F 0.47	5.5%	36953	
2	0.483	F 0.57	4.0%	27038	
3	0.727	F 0.86	7.0%	47340	
4	0.847	F 1.00	22.0%	148253	Consistent with F
5	1.017	A 0.85	5.1%	34706	3
6	1.200	A 1.01	55.6%	375631	A peak - REVIEW
7	1.385	S 0.89	0.8%	5319	

Total Area: 675240

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range