


LABORATORIO PREGEN  
Carrera 15a No 106-42  
BOGOTA

Batch 2149, Rack A, Plate 1, Well F11, 501556  
[DC6CE11C1EFA44C0] May 20, 2026 14:32:11 Pressure = 64 bar (64 to 66)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.515	F 0.59	1.8%	23516	
2	0.717	F 0.83	13.9%	182358	Acetylated F peak
3	0.848	F 0.98	77.1%	1012447	Consistent with F
4	1.218	A 1.01	7.2%	94358	A peak
Total Area:				1312679	

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range