


LABORATORIO PREGEN
 Carrera 15a No 106-42
 BOGOTA

Batch 2018, Rack A, Plate 1, Well B04, 171472
 [CD7280A15255593D] Dec 15, 2025 10:02:45 Pressure = 75 bar (74 to 77)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.400	F 0.47	1.7%	10166	
2	0.497	F 0.58	1.0%	6237	
3	0.705	F 0.83	16.2%	97337	
4	0.850	F 1.00	51.1%	307732	Consistent with F
5	1.013	F 1.19	3.5%	21179	
6	1.228	A 1.02	26.5%	159743	A peak - REVIEW
Total Area:				602394	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range