


LABORATORIO PREGEN
 Carrera 15a No 106-42
 BOGOTA

Batch 2061, Rack A, Plate 1, Well C11, 489185
 [DC739F10A9CA2642] Jan 30, 2026 11:12:18 Pressure = 90 bar (89 to 92)

FA



| PEAK | RT | REL RT | % CONC | AREA | COMMENT |
|-------------|-------|--------|--------|--------|-------------------|
| 1 | 0.403 | F 0.48 | 1.0% | 9140 | |
| 2 | 0.498 | F 0.59 | 0.5% | 4781 | |
| 3 | 0.695 | F 0.83 | 10.4% | 98359 | Acetylated F peak |
| 4 | 0.832 | F 0.99 | 48.4% | 457794 | Consistent with F |
| 5 | 0.988 | F 1.18 | 3.6% | 33584 | |
| 6 | 1.198 | A 1.01 | 36.1% | 341495 | A peak - REVIEW |
| Total Area: | | | | 945153 | |

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range