


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2080, Rack A, Plate 1, Well B12, 489570
[C9637747ABF97395] Feb 24, 2026 11:10:19 Pressure = 57 bar (57 to 59)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.413	F 0.47	1.3%	10447	
2	0.508	F 0.58	0.9%	6796	
3	0.730	F 0.83	12.8%	99029	Acetylated F peak
4	0.875	F 1.00	59.5%	461889	Consistent with F
5	1.258	A 1.01	25.5%	197992	A peak - REVIEW
Total Area:				776153	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range