


LABORATORIO PREGEN  
 Carrera 15a No 106-42  
 BOGOTA

Batch 1901, Rack A, Plate 1, Well B05, 830839  
 [88357F642691AE68] Sep 03, 2025 11:24:37 Pressure = 61 bar (61 to 63)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.407	F 0.48	2.8%	22078	
2	0.487	F 0.58	3.1%	24611	
3	0.678	F 0.81	14.8%	115612	Acetylated F peak
4	0.835	F 0.99	53.5%	417976	Consistent with F
5	0.987	F 1.17	2.8%	21688	
6	1.218	A 1.01	23.0%	179991	A peak - REVIEW
Total Area:				781956	

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range