


LABORATORIO PREGEN  
 Carrera 15a No 106-42  
 BOGOTA

Batch 1977, Rack A, Plate 1, Well D10, 859895  
 [DD37A38E68AF5B1D] Nov 14, 2025 11:42:41 Pressure = 68 bar (68 to 70)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.397	F 0.46	1.4%	14117	
2	0.493	F 0.57	0.8%	8192	
3	0.717	F 0.83	10.9%	106600	Acetylated F peak
4	0.858	F 1.00	66.6%	654697	Consistent with F
5	0.998	F 1.16	0.8%	8235	
6	1.243	A 1.01	19.4%	190527	A peak
				Total Area: 982368	

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range