


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2008, Rack A, Plate 1, Well A02, 482232
[9D72FF642E328E48] Dec 09, 2025 10:55:45 Pressure = 74 bar (73 to 76)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.400	F 0.47	1.6%	9013	
2	0.497	F 0.58	1.1%	6073	
3	0.705	F 0.83	12.3%	70681	Acetylated F peak
4	0.867	F 1.02	56.0%	320720	Consistent with F
5	1.005	F 1.18	3.1%	17820	
6	1.227	A 1.01	25.9%	148061	A peak - REVIEW
Total Area:				572368	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range