


LABORATORIO PREGEN  
Carrera 15a No 106-42  
BOGOTA

Batch 1998, Rack A, Plate 1, Well C12, 484307  
[D977BD32F9382CC8] Nov 28, 2025 12:01:13 Pressure = 71 bar (71 to 73)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.397	F 0.46	1.5%	19801	
2	0.500	F 0.58	0.5%	7212	
3	0.740	F 0.86	16.2%	216413	Acetylated F peak
4	0.850	F 0.98	72.9%	973727	Consistent with F
5	1.005	F 1.16	0.3%	4580	
6	1.242	A 1.01	8.6%	114644	A peak
Total Area:				1336377	

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range