


- - - - -
`LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2200, Rack A, Plate 1, Well B12, 487169
[8829C3389BD73915] Jul 17, 2026 15:02:27 Pressure = 82 bar (82 to 85)

FA



PEAK	RT	REL	RT	% CONC	AREA	COMMENT
1	0.683	F	0.82	22.8%	234761	
2	0.818	F	0.99	71.8%	738434	Consistent with F
3	1.162	A	1.01	5.3%	55006	A peak
Total Area: 1028201						

Codes: 1) Wide A peak
2) Area of A peak < 80%
3) Peak area greater than expected
4) Peak after A2
5) A1c > 10%
6) HbF or variant present
7) Total sample area too small/big
8) A2 is not within normal range