


LABORATORIO PREGEN
 Carrera 15a No 106-42
 BOGOTA

Batch 2147, Rack A, Plate 1, Well A10, 501858
 [9969E0D4A8706CCA] May 13, 2026 14:56:08 Pressure = 67 bar (65 to 68)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.437	F 0.23	1.4%	32560	
2	0.567	F 0.30	2.4%	55549	
3	0.878	F 0.47	11.8%	275204	Acetylated F peak
4	1.048	F 0.56	5.6%	129673	
5	1.808	F 0.97	64.6%	1508209	Consistent with F
6	2.433	F 1.30	0.2%	5455	
7	2.728	F 1.46	0.5%	12460	
8	3.068	A 0.83	0.2%	4806	
9	3.642	A 0.99	13.3%	310665	A peak
				Total Area: 2334581	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range