


LABORATORIO PREGEN
 Carrera 15a No 106-42
 BOGOTA

Batch 2049, Rack A, Plate 1, Well D09, 489154
 [C866C85F94C08CC2] Jan 20, 2026 12:30:02 Pressure = 80 bar (79 to 81)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.415	F 0.49	0.9%	9266	
2	0.503	F 0.60	0.4%	3858	
3	0.707	F 0.84	14.3%	150056	Acetylated F peak
4	0.840	F 0.99	73.2%	766647	Consistent with F
5	1.203	A 1.01	11.2%	117350	A peak
				Total Area: 1047177	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range