


LABORATORIO PREGEN  
Carrera 15a No 106-42  
BOGOTA

Batch 2061, Rack A, Plate 1, Well E06, 483336  
[9933CB50A3482642] Jan 30, 2026 11:50:42 Pressure = 89 bar (89 to 92)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.417	F 0.50	0.5%	7508	
2	0.505	F 0.60	0.5%	6587	
3	0.703	F 0.84	14.4%	206711	Acetylated F peak
4	0.837	F 1.00	70.4%	1009345	Consistent with F
5	1.187	A 1.00	14.2%	202930	A peak
Total Area:				1433081	

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range