


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2034, Rack A, Plate 1, Well C10, 171736
[CD37865654AA064A] Jan 06, 2026 11:45:20 Pressure = 83 bar (82 to 85)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.407	F 0.47	1.0%	9967	
2	0.502	F 0.59	0.7%	6999	
3	0.707	F 0.82	14.3%	143799	Acetylated F peak
4	0.847	F 0.99	62.0%	624525	Consistent with F
5	1.197	A 1.01	22.1%	222294	A peak - REVIEW

Total Area: 1007584

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range