


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2041, Rack A, Plate 1, Well E11, 171854
[8C66882306CA044A] Jan 15, 2026 12:47:28 Pressure = 83 bar (82 to 84)

FA



| PEAK | RT | REL RT | % CONC | AREA | COMMENT |
|-------------|-------|--------|--------|---------|-------------------|
| 1 | 0.413 | F 0.49 | 0.4% | 8478 | |
| 2 | 0.500 | F 0.59 | 0.4% | 8671 | |
| 3 | 0.700 | F 0.82 | 12.6% | 259969 | Acetylated F peak |
| 4 | 0.843 | F 0.99 | 64.3% | 1323863 | Consistent with F |
| 5 | 1.200 | A 1.01 | 22.2% | 456888 | A peak - REVIEW |
| Total Area: | | | | 2057869 | |

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range