


LABORATORIO PREGEN
 Carrera 15a No 106-42
 BOGOTA

Batch 1925, Rack A, Plate 1, Well B10, 478881
 [89245206DE298668] Oct 01, 2025 14:48:53 Pressure = 63 bar (61 to 64)

AF



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.393	F 0.22	4.4%	82087	
2	0.527	F 0.29	3.5%	64909	
3	0.840	F 0.47	8.5%	159575	Acetylated F peak
4	1.218	F 0.68	1.4%	26186	
5	1.782	F 0.99	32.4%	605441	Consistent with F
6	2.612	F 1.46	3.0%	56808	
7	2.837	A 0.79	1.6%	29645	
8	3.157	A 0.88	0.3%	6070	
9	3.608	A 1.00	43.8%	819359	A peak - REVIEW
10	4.180	A 1.16	0.2%	3630	
11	4.395	S 0.85	0.9%	16570	
				Total Area: 1870280	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range