


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2185, Rack A, Plate 1, Well B11, 493550
[C928BF03236DB935] Jun 25, 2026 12:31:03 Pressure = 72 bar (72 to 74)

AF



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.418	F 0.50	0.5%	8198	
2	0.507	F 0.61	0.3%	5404	
3	0.687	F 0.82	5.9%	106345	Acetylated F peak
4	0.842	F 1.01	36.0%	651024	Consistent with F
5	1.183	A 1.00	57.4%	1037747	A peak - REVIEW
Total Area:				1808718	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range