


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2080, Rack A, Plate 1, Well D06, 490061
[99233647A9737395] Feb 24, 2026 11:48:43 Pressure = 56 bar (56 to 58)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.417	F 0.47	1.2%	15319	
2	0.517	F 0.59	0.7%	8747	
3	0.730	F 0.83	18.5%	233563	Acetylated F peak
4	0.865	F 0.98	67.9%	856632	Consistent with F
5	1.022	F 1.16	0.7%	8993	
6	1.267	A 1.01	10.9%	137637	A peak
Total Area:				1260891	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range