


LABORATORIO PREGEN  
 Carrera 15a No 106-42  
 BOGOTA

Batch 2121, Rack A, Plate 1, Well F12, 498254  
 [8832125656A3F315] Apr 13, 2026 12:58:59 Pressure = 57 bar (57 to 59)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.420	F 0.48	1.2%	12781	
2	0.508	F 0.58	1.7%	18597	
3	0.715	F 0.82	13.0%	141156	Acetylated F peak
4	0.858	F 0.98	61.4%	665242	Consistent with F
5	1.003	F 1.15	2.1%	22364	
6	1.250	A 1.01	20.6%	223356	A peak - REVIEW
Total Area:				1083496	

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range